



भारतीय मानक ब्यूरो
BUREAU OF INDIAN STANDARDS

पूर्वी क्षेत्रीय प्रयोगशाला : P-230 सी.आई.टी.स्कीम VII एम,
ब्लक डब्ल्यू, वी.आई.पी.रोड, कान्कुरगछी, कोलकाता-700 054
Eastern Regional Laboratory, P-230 C.I.T. Scheme VII M,
Block-W, V.I.P. Road, Kankurgachi, Kolkata - 700 054

टेलिफोन : 2355-6586/6587 तार : मानकसंस्था
Phones : 2355-3243/3244 Grams : Manaksanstha

Website : <http://www.bis.org>
E-mail : erol@bis.org.in

By Speed Post/Courier

Our Ref: EROL/Pur (Mech)

16-March-09

Subject: Tender for supply of Humidity Cabinet (CEM 03)

TO

Dear Sirs,

Tenders are invited for the supply of the above item in the sealed cover duly superscribed as **“QUOTATION FOR THE SUPPLY OF HUMIDITY CABINET (CEM 03)”** Complying the requirements of the enclosed **Annex-A** (The Term and conditions of supply) and **Annex-B** (The Technical Specification of the equipment), which should reach the undersigned latest by **31-03-2009**. The tenders shall be opened in this office at 1500 hours in the next working day in the presence of such tenderers or their duly authorized representatives, who may like to attend. Also refer our BIS website: www.bis.org.in for any clarification.

Thanking you,

Yours faithfully,

A.Chatterjee
(Dy Director)

Encl: As above

ANNEX-A
(Terms and conditions of Supply)

- 1) The BUREAU gives FIRST PREFERENCE in its purchase to goods bearing ISI CERTIFICATION MARK and second preference to those which conform to the relevant Indian Standard Specifications.
- 2) The delivery of the equipment shall preferably be made within 30 days of receipt of order. If, however, it is not possible to affect delivery during working hours by the date, the date by which the delivery of the equipment can be effected/guaranteed may be so stated.
- 3) QUOTATIONS shall be submitted by interested Manufacturer/Supplier in the form of two BIDS/PARTS – namely, “TECHNICAL BID & FINANCIAL BID.”
- 4) Both Technical Bid & Financial Bid shall be sealed by the bidder in SEPARATE ENVELOPE/COVER duly superscribed “TECHNICAL OR FINANCIAL BID which shall finally be put in a bigger size envelope duly supercribed “QUOTATION FOR THE SUPPLY OF(Name).....)”
- 5) The rates quoted should give break up of Cost, Excise Duty/ Sales Tax/VAT, P&F charges, Freight, Insurance etc (as applicable). If the rates quoted do not include sales tax/excise duty/ or any other taxes/charges, the same should be specifically stated. **The total cost of the product must also be mentioned after summation of all the above charges.**
- 6) Tenders qualified by such value and indefinite expressions as “Subject to immediate acceptance”, “Subject to prior sale” etc., and incomplete quotation are liable to be summarily rejected.
- 7) All Tenders shall remain open for a period of 90 days from the date of opening.
- 8) Please note that Tenderer shall have to **deposit 5% of cost of equipment as EMD for bid security.** The amount is to be deposited along with the tenders, in the form of **demand draft** drawn in favor of BUREAU OF INDIAN STANDARDS payable at Kolkata. Bid security as deposited with tender shall be returned to the unsuccessful bidder after the decision is made for purchase.
- 9) Tenderer getting the order shall preferably deposit 10% of the ordered amount as contract **PERFORMANCE SECURITY** from any nationalized Bank valid till 60 days after warranty /guarantee period, which shall be returned after satisfactory performance .However ,other terms of payment such as cited at clause 10 may also be considered.
- 10) 90% payment may be made after satisfactory installation & commissioning of the equipment and **10% (contract performance security)** after expiry of guarantee/warranty period. The equipment should have a **guarantee/ warranty period of preferably 3 years & the period shall be mentioned by the Tenderer while submitting quotation .**
- 11) **The Tenders/Suppliers of the equipment shall also indicate the availability for provision of servicing / repair / maintenance of the equipment in Kolkata** and shall preferably have necessary infrastructure /capacity to provide a **comprehensive maintenance guarantee, preferably for three years** including replacement of spares, if any. Position shall be clarified in this aspect while submitting quotation.
- 12) The supplier/Manufacturer of the equipment shall ensure **training of officials** of the laboratory during commissioning of equipment.
- 13) All goods shall be received subject to approval on inspection. The decision of our Inspecting Officers shall be binding. Rejected items/goods/stores shall be removed by the supplier at his own cost and risk, within 30 days of receipt of notice for the removal of such goods, and thereafter no liability, whatsoever, on the Bureau shall be attached for the rejected/disapproved goods/items/stores.
- 14) The Bureau reserves the right to accept or reject summarily any or all tenders in whole or part without assigning any reason whatsoever.
- 15) The Bureau takes no responsibility for delay, loss or non-receipt of tenders after dispatch.
- 16) In case of non-compliance with the contract, the Bureau reserves its right to:
 - a) Cancel/rescind/revoke the order if supply is not made in time and is not Conforming to the required specification.
 - b) Impose penalty up to 1% of the total value of the order for a delay of every seven days after the schedule date subject to the ceiling of a maximum of 10% of the total value of the order.

All question, disputes or difference arising under, out of or in connection with this tender enquiry shall be subject to the exclusive jurisdiction of the Kolkata Courts.

Annex – B

HUMIDITY CABINET

QUANTITY: One No

Technical Requirements

- i) Digital display for Temperature and Humidity Controller.
- ii) Required Humidity : not less than 90% RH .
- iii) Temperature : 27 ± 2 °C.
- iv) Chamber Size (Inside): 750 x 450 x 900 mm.
- v) Least Count ± 1.0 °C and RH $\pm 1\%$ or better .
- vi) Front opening and glass door.
- vii) The material of cabinet shall be corrosion resistant.
- viii) The outer finish shall be painted/ coated and free from sharp edges.