

**ANNEXURE-I****BUREAU OF INDIAN STANDARDS  
(Southern Regional Office Laboratory)****SCHEDULE TO QUOTATION(S)**

TIME & DATE OF OPENING OF QUOTATION(S): 1500 h on 03 July 2006

The quotation(s) may remain open for acceptance for a maximum period of 120 days from the date of opening of quotation(s).

Item No.	Description of Goods / Stores	Quantity	Rate per Unit
1. Name:	<b>MICRO BALANCE</b>	1 No.	
Specification:	Capacity : 2000 mg or more Readability/Least Count: 0.0001 mg Sensitivity to temperature: $\pm 1$ ppm/ $^{\circ}$ c or better. Linearity: 0.001 mg or better		
2. Name:	<b>MICRO BALANCE</b>	1 No.	
Specification:	Capacity : 5000 mg or more Readability/Least Count: 0.001 mg Sensitivity to temperature: $\pm 1$ ppm/ $^{\circ}$ c or better. Linearity: 0.002 mg or better		
<b><u>Note:</u></b>			
i)	You may quote for the suitable balance(s) in your product range by giving your specification.		
ii)	Internal calibration facility shall be in-built in the balance.		
iii)	Quotation for the essential accessories shall be submitted		
iv)	Optional Accessories such as external calibration Weights, software for making computerized Measurement and others if any shall be quoted with their price.		

:2:

- v) Physical dimension of the balance, its weight; repeatability, overall uncertainty and stability of measurement, error due to corner loading, response time, output, power source required & operating temperature ranges, shall be stated in the quotation along with other features, if any.
- vi) After sales warranty available, service facility available in Chennai, AMC facility (with the cost) and the terms and conditions for availing these facilities shall be quoted.
- vii) Spares if any has to be kept in stock, the cost of the spares shall be quoted.